

*Full Length Research Paper*

# Evaluation and comparison of SYBR Green I Real-Time PCR and TaqMan Real-Time PCR methods for quantitative assay of *Listeria monocytogenes* in nutrient broth and milk

Hassan Dadkhah<sup>1</sup>, Mohammad Reza Bassami<sup>2</sup>, Saeed Hashemi<sup>3</sup>, Farzaneh Shahraz<sup>1</sup>, Hedayat Hosseini<sup>1</sup>, Kimon Andreas G. Karatzas<sup>4</sup> and Ramin Khaksar<sup>1\*</sup>

<sup>1</sup>Department of Food Science and Technology, National Nutrition and Food Technology Research Institute, Faculty of Nutrition Science and Food Technology, Shaheed Beheshti University, M.C., Tehran, Iran.

<sup>2</sup>Institute of Biotechnology and Faculty of Veterinary Medicine, Ferdowsi University of Mashhad, Iran.

<sup>3</sup>Endocrinology and Metabolism Research Centre, Tehran University and Medical Sciences, Dr Shariati Hospital, North Kargar St., 14114 Tehran, Iran.

<sup>4</sup>Department of Microbiology, National University of Ireland, Galway, Galway, Ireland.

Accepted 27 December, 2011

Specific traditional plate count method and real-time PCR systems based on SYBR Green I and TaqMan technologies using a specific primer pair and probe for amplification of *iap*-gene were used for quantitative assay of *Listeria monocytogenes* in seven decimal serial dilution series of nutrient broth and milk samples containing 1.58 to 1.58×10<sup>7</sup> cfu/ml and the real-time PCR methods were compared with the plate count method with respect to accuracy and sensitivity. In this study, the plate count method was performed using surface-plating of 0.1 ml of each sample on Palcam Agar. The lowest detectable level for this method was 1.58×10<sup>3</sup> cfu/ml for both nutrient broth and milk samples. Using purified DNA as a template for generation of standard curves, as few as four copies of the *iap*-gene could be detected per reaction with both real-time PCR assays, indicating that they were highly sensitive. When these real-time PCR assays were applied to quantification of *L. monocytogenes* in decimal serial dilution series of nutrient broth and milk samples, 3.16×10<sup>3</sup> to 3.16×10<sup>5</sup> copies per reaction (equals to 1.58×10<sup>3</sup> to 1.58×10<sup>7</sup> cfu/ml *L. monocytogenes*) were detectable. As logarithmic cycles, for Plate Count and both molecular assays, the quantitative results of the detectable steps were similar to the inoculation levels.

**Key words:** *Listeria monocytogenes*, real-time PCR, milk, pathogen.

## INTRODUCTION

*Listeria monocytogenes* is a Gram-positive, anaerobic and facultative intracellular bacterium that is recognized worldwide as one of the most important food-borne pathogens of concern for the food industries (Gray et al., 2004; Nakamura et al., 2004; Rodríguez-Lázaro et al., 2004b; Cocolin et al., 2005; Berrada et al., 2006;

Rossmannith et al., 2006; Zhang et al., 2007; Rantsiou et al., 2008). This micro-organism has been responsible for listeriosis outbreaks in the past years (Amagliani et al., 2004; Cocolin et al., 2005; Ueda et al., 2005; Kim et al., 2006) and can be isolated from foods of animal origin such as milk and milk products (Nogva et al., 2000; Kačlíková et al., 2004; Makino et al., 2005; Kalorey et al., 2008; Rantsiou et al., 2008). The presence of *L. monocytogenes* in milk-based products can be related to raw milk contamination or to post-pasteurization contamination (Carminati et al., 2004). Furthermore,

\*Corresponding author. E-mail: [r.khaksar@sbm.ac.ir](mailto:r.khaksar@sbm.ac.ir). Tel/Fax: (+98 21) 22376426.

several outbreaks of listeriosis have been proven to be associated with the consumption of milk and are a great concern for the dairy industry (Amagliani et al., 2004). Contamination of milk with this pathogen therefore poses a great health risk to humans.

The detection methods for food borne pathogens generally involve colony isolation on selective media and the use of biochemical tests and serotyping with antibodies against specific bacterial antigens (Aslam et al., 2003; Flekna et al., 2007; Badosa et al., 2009; D'Urso et al., 2009). These conventional methods are laborious and time consuming, and several days are therefore required before results are obtained (Amita et al., 2002; Cocolin et al., 2002; Berrada et al., 2006; Rodríguez-Lázaro and Hernández, 2006; Oravcová et al., 2007; Reichert-Schwillinsky et al., 2009). Thus, the development of rapid, sensitive and species-specific methods for the detection and quantification of *L. monocytogenes* in milk are required to overcome these limitations.

In the last 10 years, Real-Time PCR systems based on SYBR Green I and TaqMan technologies have increasingly been used for accurate and reliable detection and quantification of various food-borne pathogens (Rodríguez-Lázaro et al., 2004a; Rawsthorne and Phister, 2006; Skånseng et al., 2006; Enokimoto et al., 2007; Malorny et al., 2007), including *L. monocytogenes* in contaminated samples along the food production chain. Compared to conventional PCR-based methods, Real-Time PCR technologies involve a lower risk of cross-contamination because the presence of the target sequence(s) in the sample is indicated by an increase in fluorescence signal, and no post-PCR processing of the sample is required (Rodríguez-Lázaro et al., 2004c; Rossmannith et al., 2006). The detection of *Listeria monocytogenes* was accomplished by Real-Time PCR methods directed to the *ssrA* (O'Grady et al., 2008), the *hly* (Norton et al., 2000; Rodríguez-Lázaro et al., 2005), the *actA* (Norton et al., 2001; Oravcová et al., 2006), the *prfA* (Rossmannith et al., 2006), the *iap* (Comi et al., 1997; Hein et al., 2001) and the *inl* (Almeida and Almeida, 2000; Pangallo et al., 2001) genes. Each of these methods is rapid, sensitive and has proven to be specific, but most of them have not yet been compared to the Plate Count method with respect to sensitivity, precision (determined by the standard deviation among replicates of one sample) and accuracy.

Moreover, in Real-Time PCR-based detection systems, amplification efficiencies can be different from sample to sample (Klerks et al., 2004) and application of these systems for direct quantification of pathogens present in foods has been limited by the complex composition of the starting materials which contain inhibitors for PCR amplification. For example, when PCR was applied to milk samples, its sensitivity was low when compared to that of bacterial cultures (Amagliani et al., 2004), particularly at low concentrations of the pathogen (Amagliani et al., 2004).

The aim of this study was to evaluate and compare the

Plate Count method against a SYBR Green I and TaqMan-based Real-Time PCR method for quantitative assay of *L. monocytogenes* in artificially contaminated nutrient broth and milk samples. The applicability of both methods was determined with respect to accuracy and sensitivity.

## MATERIALS AND METHODS

### Bacterial strain, medium, and culture conditions

*L. monocytogenes* ATCC 19115 was obtained from the Iranian Research Organization for Science and Technology (IROST) and used in all molecular and culture methods. The strain was grown overnight at 37°C in Nutrient Broth (Merck, Germany). Subsequently, using a culture containing  $1.58 \times 10^8$  cfu *L. monocytogenes* per ml, decimal serial dilutions containing  $1.58 \times 10^7$  cfu/ml were prepared in the nutrient broth itself. The cell concentrations of these decimal serial dilutions were estimated in triplicate by surface-plating of 0.1 ml of each sample on Palcam agar (Merck, KGaA, Darmstadt, Germany) following incubation at 37°C for 24 h. For Real-Time PCR quantification, the serial dilutions were directly subjected to DNA extraction. Moreover, *E. coli* O157:H7 ATCC 43895 (IROST) was used as negative control in molecular methods.

### Artificially contaminated milk samples

UHT-treated milk (2.5% fat) was purchased from a local supermarket. Subsequently, decimal serial dilutions of milk containing 1.58 to  $1.58 \times 10^7$  cfu *L. monocytogenes* per ml were prepared. These serial dilutions were prepared by the addition of 1 ml of each of the decimal serial dilutions of nutrient broth containing  $1.58 \times 10^8$  cfu *L. monocytogenes* per ml to 9 ml of UHT-treated milk. Similar to contaminated nutrient broth samples, the bacterial levels in each contaminated milk sample were determined in triplicate by surface-plating of 0.1 ml of each sample on PALCAM Agar (Merck, Germany) followed by incubation at 37°C for 24 h. After the artificial contamination of the milk samples, for molecular quantification of the bacterial numbers, the serial dilutions were directly subjected to DNA extraction.

### DNA extraction

DNA extraction from both cultures and contaminated milk samples was performed as follows. Five hundred  $\mu$ l of TE buffer (10mM Tris, 1 mM EDTA, pH 8.0) was added to 500 $\mu$ l of each sample, incubated at 85-90°C for 15 min and immediately placed at -20°C for 20 min. Once samples were removed from the freezer, 100 $\mu$ l (20 mg/ml) lysozyme was added and incubated at 37°C for 2 hrs. Subsequently, 75 $\mu$ l proteinase K (20 mg/ml) was added followed by incubation at 60-65°C for 20 min.

Then, equal parts of 470 $\mu$ l of lysis buffer, isopropanol and each sample were transferred to a 1.5 ml sterile filter tube and centrifuged (10,000  $\times$  g for 1 min). The supernatant was washed once with 750  $\mu$ l of wash buffer I and twice with 750  $\mu$ l of wash buffer II (Roche Molecular Biochemicals, Indianapolis, IN, USA). Following centrifugation at 14,000  $\times$  g for 3 min, 50  $\mu$ l of pre-warm elution buffer (Roche Molecular Biochemicals, Indianapolis, IN, USA) was added. After incubation for 5 min. at room temperature and final centrifugation step (8000  $\times$  g for 1 min), 5 $\mu$ l of the final elute was used for quantitative Real-Time PCR reactions with respect to factor of 50.

**Table 1.** Sequences of the oligonucleotides and their locations in the *iap*-gene of *L. monocytogenes*.

Primer name	Sequence (5' to 3')	Location	Concentration in the PCR mix (nM)	Reference
LIM 2 (forward primer)	CTA AAG CGG GAA TCT CCC TT	1214–1233	250	Hein et al., 2001
LIMRE (reverse primer)	CCA TTG TCT TGC GCG TTA AT	1369–1388	250	Hein et al., 2001
LMS 2 (probe)	FAM-CTT CTG GCG CAC AAT ACG CTA GCA CT-TAMRA	1241–1266	250	Hein et al., 2001

### Primers and probe

Detection of *L. monocytogenes* was based on the amplification of *iap* gene with the use of PCR-primers and probe (Table 1) as previously described by Hein et al. (2001). Quantification using SYBR Green I required the use of the LIM 2 and LIMRE primers, while the use of TaqMan required the use of the previous primer pair in combination with LMS 2 probe. The specific TaqMan probe was labelled at the 5' and the 3' end with a 6-carboxy-fluorescein group (FAM) and 6-carboxytetramethyl-rhodamine (TAMRA), respectively. Primers and probe were synthesized by Metabion (Munich, Germany).

### Real-Time PCR amplification conditions

For generation of standard curves and quantitative assays, amplifications were carried out with 5µl of elute, extracted as described above, in a final volume of 20µl using the MiniOpticon Real-Time PCR Detection System (BioRad, Hercules, CA, USA).

For detection with SYBR Green I, following the addition of 5µl of elute in the tube, 4.0µl of 5x Eva Green (BioAtlas, Estonia and Finland) was added. MgCl<sub>2</sub> was also added up to a concentration of 5mM while primers were added up to a final concentration of 250nM. Based on optimization of annealing temperature, the cycling conditions included an initial denaturation at 95°C for 15 min followed by 45 cycles of 15 s denaturation at 95°C, 30 s annealing at 54°C and 30 s extension at 72°C. The thermocycling program was followed by a melting program of 95°C for 1 min. (denaturation), 45°C for 30 s (annealing), and then a transition from 45 to 95°C with a rate of 0.1°C/s. Monitoring of fluorescence was performed at regular intervals during the annealing step and continuously throughout the melting phase.

For detection with TaqMan, following the addition 5µl elute of extracted DNA to the tube 4.0 µl of 5xAmpliTaq (BioAtlas, Estonia and Finland) was added. MgCl<sub>2</sub> was also added up to a concentration of 5mM while primers and probe were added at a final concentration of 250 nM. The amplification proceeded with an initial step of 95°C for 15 min followed by 45 cycles of 95°C for 15 s and 57°C for 1 min. (temperature optimized through a gradient protocol). All molecular assays were performed in triplicate for each sample.

### Generation of standard curves

DNA extraction from *L. monocytogenes* used for construction of standard curves and sensitivity testing was carried out using a High Pure PCR template preparation kit (Roche Molecular Biochemicals, Indianapolis, IN, USA) as described by the manufacturer.

The concentration of the extracted DNA was measured by NanoDrop ND-1000 Spectrophotometer (BioPhotometer, Eppendorf, Hamburg, Germany). The correlation of the signal detected by the increase of the amplicons during the Real-Time PCR to the original number of copies of the *iap* gene was based on

the fact that only a single copy of the *iap*-gene is contained in the genome (Hein et al., 2001) and that 1ng of listerial DNA equals 3.1×10<sup>5</sup> copies of the whole genome (Rossmannith et al., 2006).

## RESULTS


### Standard curves and linear range of quantification

The concentration of the extracted DNA used for construction of standard curves was 0.258 µg/µl, which corresponds to 8×10<sup>7</sup> copies of the *iap*-gene. To estimate the detection limits of the Real-Time PCR assays, 6 serial decimal dilutions of the extracted DNA containing 4×10<sup>5</sup> to 4 copies of the *iap*-gene in the 5 µl used for each reaction were prepared. Both Real-Time PCR methods were very sensitive and they were able to detect as few as four copies per reaction. Standard curves of the logarithm of the quantity of the template against the C<sub>T</sub> were constructed (Figure 1) and their slopes with SYBR Green I and TaqMan were -3.395 and -3.230 respectively while their coefficients of correlation (r<sup>2</sup>) were 0.999 and 1, respectively, indicating that the Real-Time PCR systems were highly linear over a range of five logarithms.

The specificity of both Real-Time PCR assays was assessed by examining 10-fold serial dilutions of whole genomic DNA of *L. monocytogenes*. For both molecular assays, agarose gel (1.5%) electrophoresis of the amplification products of these dilution series were visualized under UV light and all displayed the expected band of 175 bp. Moreover, no amplified DNA fragments were observed when the primers and probe from two negative controls (a different organism's DNA and a control with no DNA template) (Figure 2).

### Quantitative detection of artificially contaminated nutrient broth and milk samples using plate count method

Nutrient broth and milk samples artificially contaminated by the serial decimal dilution series of *L. monocytogenes* ranging from 1.58 to 1.58×10<sup>7</sup> cfu/ml were analyzed in triplicate by surface-plating on PALCAM Agar (Table 2). The plate count method was able to detect the numbers estimated according to the inoculation levels. The lowest detectable level of *L. monocytogenes* was 1.58×10 cfu



**Figure 1.** Standard curves of the SYBR-Green I Real-Time PCR (A) and TaqMan Real-Time PCR (B) performed using as template the 10-fold serial dilutions of genomic DNA extracted from *L. monocytogenes*.

per ml nutrient broth and milk samples. The mean of the three replicates measurements of the lowest detectable level by Plate Count method was  $1.33 \times 10$  and  $2.00 \times 10$  cfu per ml nutrient broth and milk samples, respectively.


#### Calculation of the copy number of the *iap*-gene in artificially contaminated nutrient broth and milk samples using SYBR Green I Real-Time PCR

Table 3 shows the results of SYBR Green I detection of artificially contaminated nutrient broth and milk samples. For both inoculated nutrient broth and milk samples, five decimal dilution series of the extracted genomic DNA, ranged from  $3.16 \times 10$  to  $3.16 \times 10^5$  copies per reaction were detectable. These five dilutions were extracted from inoculated nutrient broth and milk samples containing  $1.58 \times 10^3$  to  $1.58 \times 10^7$  cfu/ml *L. monocytogenes*. Therefore, using SYBR Green I molecular assay, the lowest detectable level of genomic DNA extracted from nutrient broth and milk samples was  $3.16 \times 10$  copies per

reaction, which equals  $1.58 \times 10^3$  cfu *L. monocytogenes* per ml. The mean of the three replicates measurements of this level by SYBR Green I system was  $2.4 \times 10$  copies per reaction (equal to  $1.20 \times 10^3$  cfu/ml) for nutrient broth samples and  $4.22 \times 10$  copies per reaction (equal to  $2.11 \times 10^3$  cfu/ml) for milk samples. Moreover, for both nutrient broth and milk samples the five detectable dilutions showed the numbers estimated according to logarithmic decades of the inoculation levels.

#### Calculation of the copy number of the *iap*-gene in artificially contaminated nutrient broth and milk samples using TaqMan Real-Time PCR

Similar to those obtained for SYBR Green I quantitative assay, five decimal dilution series of the extracted genomic DNA, ranging from  $3.16 \times 10$  to  $3.16 \times 10^5$  copies per reaction, were detectable using TaqMan molecular assay (Table 4). As SYBR Green I system, these five dilutions were extracted from inoculated nutrient broth



**Figure 2.** Analytical sensitivity results of amplification products of the SYBR-Green I Real-Time PCR (A) and TaqMan Real-Time PCR (B). Lane 1: 100-bp ladder; lane 2: control with no DNA template; lanes 3–8: ten-fold dilution series of whole genomic DNA (ranging from  $4 \times 10^5$  to  $4 \times 10^0$  copies of the *iap* gene/PCR reaction); lane 9: different organism's DNA; lane 10: 100-bp ladder.

and milk samples containing  $1.58 \times 10^3$  to  $1.58 \times 10^7$  cfu/ml *L. monocytogenes*. Moreover, the obtained quantitative results of evaluation of these five detectable steps showed the same logarithmic decades of the inoculation levels. The mean of the three replicated measurements of the lowest detectable level ( $3.16 \times 10$  copies per reaction, equal to  $1.58 \times 10^3$  cfu/ml *L. monocytogenes*) by TaqMan quantitative assay was  $2.38 \times 10$  copies per reaction (equal to  $1.19 \times 10^3$  cfu/ml) for nutrient broth samples and  $2.86 \times 10$  copies per reaction (equal to  $1.43 \times 10^3$  cfu/ml) for milk samples.

## DISCUSSION

PCR is a technique that enables the detection of the presence of food-borne pathogens. Several papers have focused on its use for the detection and quantification of *L. monocytogenes*. However, few studies are available which compare PCR-based assays with traditional Plate Count methods with respect to their accuracy and sensitivity. Hence, in this study, using a specific primer pair and probe, targeting the *iap*-gene, two Real-Time PCR methods (SYBR Green I and TaqMan Real-Time PCR assays) were used for specific quantification of decimal serial dilutions of *L. monocytogenes* in artificially

contaminated nutrient broth and milk samples. Finally, after evaluation of the obtained results of these quantitative assays the results were compared with the obtained results of traditional Plate Count method using PALCAM Agar, with respect to accuracy and sensitivity.

The Plate Count method showed quantitative results of expected logarithmic decades, indicating the accuracy of this method for quantification detection of artificially contaminated nutrient broth and milk samples. Because of the plating of 0.1 ml of each sample the samples inoculated by cfu < 10 per ml sample were not detectable. Using purified DNA dilutions as a template, as few as four *iap*-gene copy numbers per reaction were detected by the both molecular assays, which was two copy numbers lower than that reported by Hein et al. (2001).

The numbers of cells estimated with both Real-Time PCR assays were similar to those obtained by the plate count method, indicating the accuracy of these methods for the quantification of *L. monocytogenes* present in artificially contaminated nutrient broth and milk samples. The dilution coefficient of 50 in DNA extraction step made the samples inoculated by cfu < 50 per ml not detectable. However, because of PCR inhibitors and other factors, the samples inoculated by  $1.58 \times 10^2$  cfu/ml were not detectable. Quantitative results of SYBR Green I and TaqMan systems indicate the seam sensitivity and

**Table 2.** Quantitative detection of 10-fold serial dilutions of *L. monocytogenes* in artificially contaminated nutrient broth and milk samples using Plate Count method. The samples were evaluated in triplicate.

Sample	Plate count method (CFU/ml)×10					
	Nutrient broth	Mean	SD	Milk	Mean	SD
1	1.83×10 <sup>7</sup>	1.66×10 <sup>7</sup>	0.32×10 <sup>7</sup>	1.96×10 <sup>7</sup>	2.01×10 <sup>7</sup>	0.19×10 <sup>7</sup>
	1.29×10 <sup>7</sup>			2.22×10 <sup>7</sup>		
	1.86×10 <sup>7</sup>			1.85×10 <sup>7</sup>		
2	1.15×10 <sup>6</sup>	1.15×10 <sup>6</sup>	0.09×10 <sup>6</sup>	2.84×10 <sup>6</sup>	1.93×10 <sup>6</sup>	0.81×10 <sup>6</sup>
	1.06×10 <sup>6</sup>			1.28×10 <sup>6</sup>		
	1.24×10 <sup>6</sup>			1.67×10 <sup>6</sup>		
3	2.52×10 <sup>5</sup>	2.11×10 <sup>5</sup>	0.36×10 <sup>5</sup>	3.21×10 <sup>5</sup>	2.54×10 <sup>5</sup>	1.29×10 <sup>5</sup>
	2.00×10 <sup>5</sup>			1.05×10 <sup>5</sup>		
	1.81×10 <sup>5</sup>			3.37×10 <sup>5</sup>		
4	1.16×10 <sup>4</sup>	2.51×10 <sup>4</sup>	1.17×10 <sup>4</sup>	1.31×10 <sup>4</sup>	1.40×10 <sup>4</sup>	0.47×10 <sup>4</sup>
	3.04×10 <sup>4</sup>			1.91×10 <sup>4</sup>		
	3.33×10 <sup>4</sup>			0.98×10 <sup>4</sup>		
5	2.44×10 <sup>3</sup>	2.24×10 <sup>3</sup>	0.88×10 <sup>3</sup>	1.59×10 <sup>3</sup>	1.76×10 <sup>3</sup>	0.37×10 <sup>3</sup>
	1.27×10 <sup>3</sup>			1.50×10 <sup>3</sup>		
	3.00×10 <sup>3</sup>			2.19×10 <sup>3</sup>		
6	1.90×10 <sup>2</sup>	1.60×10 <sup>2</sup>	0.36×10 <sup>2</sup>	1×10 <sup>2</sup>	1.20×10 <sup>2</sup>	0.26×10 <sup>2</sup>
	1.20×10 <sup>2</sup>			1.10×10 <sup>2</sup>		
	1.70×10 <sup>2</sup>			1.50×10 <sup>2</sup>		
7	1×10	1.33×10	0.57×10	4×10	2.00×10	1.73×10
	2×10			1×10		
	1×10			1×10		
8	ND	ND	-	ND	0.33	-
	ND			ND		
	ND			ND		

ND: Not detectable.

accuracy of these methods. However, due to the high cost of probe synthesis, the TaqMan Real-Time PCR system cannot necessarily be used for quantification and detection of *L. monocytogenes*.

Generally, the potential of the investigated techniques for quantification of *L. monocytogenes* are influenced by a number of factors that are not completely understood. Depending on the distribution of the microorganisms in the sample and the presence of the aggregated bacterial cells, different quantitative results may be obtained by the Plate Count method. In addition, the sample preparation step prior to detection is of key importance when applying Plate Count or molecular systems to detection and quantification of microorganisms in foods. If the levels of contamination are low and small volumes of the sample are used in the plating or extraction steps, or a high

dilution of coefficients is used in these steps, the bacterial cells might not be detected and quantified. Sometimes dead and viable but non-culturable bacteria cells that are unable to grow on solid agar media, and therefore cannot be quantified by the Plate Count method, are also present. However, these cells can be detected and quantified using the Real-Time PCR methods. This poses a serious problem since the main purpose of the detection of *L. monocytogenes* in foods is the food safety and the protection of the consumer from a potentially hazardous bacterium, and it is obvious that dead or viable but non-culturable cells do not pose any threat to the consumer. Therefore, it is possible that the implementation of these Real-Time PCR methods on their own might lead to high fines and false rejection of absolutely safe products, which can lead to unfair treatment of food

**Table 3.** Quantitative detection of 10-fold serial dilutions of *iap* gene copy numbers in artificially contaminated nutrient broth and milk samples using SYBR-Green I Real-Time PCR assay. Dependently DNA extraction was used for the three replicate of each sample.

Sample	SYBR-Green I Real-Time PCR assay					
	Nutrient broth			Milk		
	Copy number/ml×50	Mean	SD	Copy number/ml×50	Mean	SD
1	2.00×10 <sup>7</sup>	1.33×10 <sup>7</sup>	0.57×10 <sup>7</sup>	1.06×10 <sup>7</sup>	1.11×10 <sup>7</sup>	0.09×10 <sup>7</sup>
	1.00×10 <sup>7</sup>			1.06×10 <sup>7</sup>		
	1.00×10 <sup>7</sup>			1.22×10 <sup>7</sup>		
2	1.01×10 <sup>6</sup>	1.25×10 <sup>6</sup>	0.33×10 <sup>6</sup>	3.60×10 <sup>6</sup>	2.73×10 <sup>6</sup>	0.75×10 <sup>6</sup>
	1.63×10 <sup>6</sup>			2.39×10 <sup>6</sup>		
	1.11×10 <sup>6</sup>			2.21×10 <sup>6</sup>		
3	3.00×10 <sup>5</sup>	3.00×10 <sup>5</sup>	0.00×10 <sup>5</sup>	1.50×10 <sup>5</sup>	1.50×10 <sup>5</sup>	0.00×10 <sup>5</sup>
	3.00×10 <sup>5</sup>			1.50×10 <sup>5</sup>		
	3.00×10 <sup>5</sup>			1.50×10 <sup>5</sup>		
4	0.80×10 <sup>4</sup>	0.90×10 <sup>4</sup>	0.26×10 <sup>4</sup>	1.60×10 <sup>4</sup>	1.69×10 <sup>4</sup>	0.09×10 <sup>4</sup>
	0.71×10 <sup>4</sup>			1.69×10 <sup>4</sup>		
	1.20×10 <sup>4</sup>			1.78×10 <sup>4</sup>		
5	0.91×10 <sup>3</sup>	1.20×10 <sup>3</sup>	0.42×10 <sup>3</sup>	2.23×10 <sup>3</sup>	2.11×10 <sup>3</sup>	0.12×10 <sup>3</sup>
	1.69×10 <sup>3</sup>			1.99×10 <sup>3</sup>		
	1.00×10 <sup>3</sup>			2.11×10 <sup>3</sup>		
6	ND	ND	-	ND	ND	-
	ND			ND		
	ND			ND		
7	ND	ND	-	ND	ND	-
	ND			ND		
	ND			ND		
8	ND	ND	-	ND	ND	-
	ND			ND		
	ND			ND		

ND: Not detectable.

producers and high economic costs. In PCR-based assays the efficiency of amplification is influenced by the presence of "PCR inhibitors". These matrix-associated interfering substances may be attributed to the food sample itself or the culture medium. Many of "PCR inhibitors" have not been identified but some are known substances. These substances include chelators of cations and substances that degrade or bind the polymerase or the DNA template. For example, milk contains high levels of cations (Ca<sup>++</sup>), nucleases, fatty acids, proteases and DNA. Therefore, an extraction protocol for removing inhibitors similar to the one described here could reduce the concentration of PCR inhibitors, for DNA extraction from milk and nutrient broth

samples. In addition, the application of an efficiency test of decimal serial dilutions could show the presence of these inhibitors. If these problems can be overcome, these Real-Time PCR-based assays, due their accuracy and rapidity, offer a good alternative to the Plate Count method.

Detection and quantification of *L. monocytogenes* and other food-borne pathogenic microorganisms in milk as a very important food product have played a significant role in the control and prevention of further human infections. Therefore, numerous studies have been reported in the literature on the development of rapid, specific and sensitive molecular methods based on new target genes and optimized conditions. According to these published

**Table 4.** Quantitative detection of 10-fold serial dilutions of *iap* gene copy numbers in artificially contaminated nutrient broth and milk samples using TaqMan Real-Time PCR assay. Dependently DNA extraction was used for the three replicate of each sample.

Sample	TaqMan Real-Time PCR assay					
	nutrient broth			Milk		
	Copy number/ml×50	Mean	SD	Copy number/ml×50	Mean	SD
1	1.59×10 <sup>7</sup>	1.40×10 <sup>7</sup>	0.26×10 <sup>7</sup>	2.07×10 <sup>7</sup>	2.07×10 <sup>7</sup>	0.00×10 <sup>7</sup>
	1.50×10 <sup>7</sup>			2.07×10 <sup>7</sup>		
	1.09×10 <sup>7</sup>			2.07×10 <sup>7</sup>		
2	1.00×10 <sup>6</sup>	1.83×10 <sup>6</sup>	1.44×10 <sup>6</sup>	0.75×10 <sup>6</sup>	1.00×10 <sup>6</sup>	0.21×10 <sup>6</sup>
	3.50×10 <sup>6</sup>			1.13×10 <sup>6</sup>		
	1.00×10 <sup>6</sup>			1.13×10 <sup>6</sup>		
3	1.63×10 <sup>5</sup>	1.30×10 <sup>5</sup>	0.54×10 <sup>5</sup>	1.70×10 <sup>5</sup>	1.60×10 <sup>5</sup>	0.13×10 <sup>5</sup>
	1.60×10 <sup>5</sup>			1.65×10 <sup>5</sup>		
	0.67×10 <sup>5</sup>			1.45×10 <sup>5</sup>		
4	2.00×10 <sup>4</sup>	2.13×10 <sup>4</sup>	0.11×10 <sup>4</sup>	1.04×10 <sup>4</sup>	1.04×10 <sup>4</sup>	0.00×10 <sup>4</sup>
	2.20×10 <sup>4</sup>			1.04×10 <sup>4</sup>		
	2.20×10 <sup>4</sup>			1.04×10 <sup>4</sup>		
5	1.19×10 <sup>3</sup>	1.19×10 <sup>3</sup>	0.00×10 <sup>3</sup>	1.11×10 <sup>3</sup>	1.43×10 <sup>3</sup>	0.42×10 <sup>3</sup>
	1.19×10 <sup>3</sup>			1.91×10 <sup>3</sup>		
	1.19×10 <sup>3</sup>			1.28×10 <sup>3</sup>		
6	ND	ND	-	ND	ND	-
	ND			ND		
	ND			ND		
7	ND	ND	-	ND	ND	-
	ND			ND		
	ND			ND		
8	ND	ND	-	ND	ND	-
	ND			ND		
	ND			ND		

ND: Not detectable.

molecular protocols, development of specific target genes, primer pairs, probe and other conditions seem more than sufficient for specific identification of *L. monocytogenes* in different foods, such as milk and milk products. Therefore, future studies must be focused mainly on efficient DNA extraction from different matrixes, development of new protocols for specific identification of other food-borne pathogenic microorganisms in food products and validation of these alternative methods with respect to sensitivity, specificity and accuracy.

## Conclusions

In this study, we evaluated and compared two Real-Time PCR quantitative methods against the Plate Count

method for quantification of dilution series of *L. monocytogenes* in two different matrixes. The Real-Time PCR methods showed similar accuracy for quantitative detection of examined samples, but the sensitivity of Plate Count method was two logs lower than the investigated molecular assays. Finally, according to the obtained results and with respect to the advantages of the molecular systems, these assays can be considered a powerful alternative to traditional cultural methods of pathogen quantification in different foods and culture media matrixes.

## REFERENCES

Almeida PF, Almeida RCC (2000). A PCR protocol using *inl* gene as a target for specific detection of *Listeria monocytogenes*. Food Control.,



11: 97-101.

- Amagliani G, Brandi G, Omiccioli E, Casiere A, Bruce IJ, Magnani M (2004). Direct detection of *Listeria monocytogenes* from milk by magnetic based DNA isolation and PCR. *Food Microbiol.*, 21: 597-603.
- Amita J, Vandana T, Guleria RS, Verma RK (2002). Qualitative evaluation of mycobacterial DNA extraction protocols for polymerase chain reaction. *Mol. Biol. Today.*, 3: 43-49.
- Aslam M, Hogan J, Smith KL (2003). Development of a PCR-based assay to detect Shiga toxin-producing *Escherichia coli*, *Listeria monocytogenes*, and *Salmonella* in milk. *Food Microbiol.*, 20: 345-350.
- Badosa E, Chico N, Pla M, Parés D, Montesinos E (2009). Evaluation of ISO enrichment real-time PCR methods with internal amplification control for detection of *Listeria monocytogenes* and *Salmonella enterica* in fresh fruit and vegetables. *Lett. Appl. Microbiol.*, 49: 105-111.
- Berrada H, Soriano JM, Picó Y, Mañes J (2006). Quantification of *Listeria monocytogenes* in salads by real time quantitative PCR. *Intl. J. Food Microbiol.*, 107: 202-206.
- Carminati D, Perrone A, Giraffa G, Neviani E, Mucchetti G (2004). Characterization of *Listeria monocytogenes* strains isolated from Gorgonzola cheese rinds. *Food Microbiol.*, 21: 801-807.
- Cocolin L, Rantsiou K, Iacumin L, Cantoni C, Comi G (2002). Direct identification in food samples of *Listeria* spp. and *Listeria monocytogenes* by molecular methods. *Appl. Environ. Microbiol.*, 68: 6273-6282.
- Cocolin L, Stella S, Nappi R, Bozzetta E, Cantoni C, Comi G (2005). Analysis of PCR-based methods for characterization of *Listeria monocytogenes* strains isolated from different sources. *Intl. J. Food Microbiol.*, 103: 167-178.
- Comi G, Cocolin L, Cantoni C, Manzano M (1997). A RE-PCR method to distinguish *Listeria monocytogenes* serovars. *FEMS Immunol Med Microbiol.*, 18: 99-104.
- D'Urso OF, Poltronieri P, Marsigliante S, Storelli C, Hernández M, Rodríguez-Lázaro D (2009). A filtration-based real-time PCR method for the quantitative detection of viable *Salmonella enterica* and *Listeria monocytogenes* in food samples. *Food Microbiol.*, 26: 311-316.
- Enokimoto M, Kubo M, Bozono Y, Mieno Y, Misawa N (2007). Enumeration and identification of *Campylobacter* species in the liver and bile of slaughtered cattle. *Intl. J. Food Microbiol.*, 118: 259-263.
- Flekna G, Štefanič P, Wagner M, Smulders FJM, Možina SS, Hein I (2007). Insufficient differentiation of live and dead *Campylobacter jejuni* and *Listeria monocytogenes* cells by ethidium monoazide (EMA) compromises EMA/real-time PCR. *Res. Microbiol.*, 158: 405-412.
- Gray MJ, Zadoks RN, Fortes ED, Dogan B, Cai S, Chen Y, Scott VN, Gombas DE, Boor KJ, Wiedmann M (2004). *Listeria monocytogenes* isolates from foods and humans form distinct but overlapping populations. *Appl. Environ. Microbiol.*, 70: 5833-5841.
- Hein I, Klein D, Lehner A, Bubert A, Brandl E, Wagner M (2001). Detection and quantification of the *iap* gene of *Listeria monocytogenes* and *Listeria innocua* by a new real-time quantitative PCR assay. *Res. Microbiol.*, 152: 37-46.
- Kaclíková E, Oravcová K, Kuchta T, Pangallo D (2004). Comparison of three polymerase chain reaction-based methods for the detection of *Listeria monocytogenes* in food. *J Rapid Meth Autom Microbiol.*, 12: 107-113.
- Kalorey DR, Warke SR, Kurkure NV, Rawool DB, Barbudhe SB (2008). *Listeria* species in bovine raw milk: A large survey of Central India. *Food Control.*, 19: 109-112.
- Kim J, Demeke T, Clear RM, Patrick SK (2006). Simultaneous detection by PCR of *Escherichia coli*, *Listeria monocytogenes* and *Salmonella typhimurium* in artificially inoculated wheat grain. *Intl. J. Food Microbiol.*, 111: 21-25.
- Klerks MM, Zijlstra C, Van Bruggen AHC (2004). Comparison of real-time PCR methods for detection of *Salmonella enterica* and *Escherichia coli* O157:H7, and introduction of a general internal amplification control. *J. Microbiol. Methods*, 59: 337-349.
- Makino SI, Kawamoto K, Takeshi K, Okada Y, Yamasaki M, Yamamoto S, Igimi S (2005). An outbreak of food-borne listeriosis due to cheese in Japan, during 2001. *Intl. J. Food Microbiol.*, 104: 189-196.
- Malorny B, Bunge C, Helmuth R (2007). A real-time PCR for the detection of *Salmonella Enteritidis* in poultry meat and consumption eggs. *J. Microbiol. Methods.*, 70: 245-251.
- Nakamura H, Hatanaka M, Ochi K, Nagao M, Ogasawara J, Hase A, Kitase T, Haruki K, Nishikawa Y (2004). *Listeria monocytogenes* isolated from cold-smoked fish products in Osaka City, Japan. *Intl. J. Food Microbiol.*, 94: 323-328.
- Nogva HK, Rudi K, Naterstad K, Holck A, Lillehaug D (2000). Application of 5'-nuclease PCR for quantitative detection of *Listeria monocytogenes* in pure cultures, water, skim milk, and unpasteurized whole milk. *Appl. Environ. Microbiol.*, 66: 4266-4271.
- Norton DM, McCamey M, Boor KJ, Wiedmann M (2000). Application of the BAX for screening/genus *Listeria* polymerase chain reaction system for monitoring *Listeria* species in cold-smoked fish and in the smoked fish processing environment. *J. Food Prot.*, 63: 343-346.
- Norton DM, Scarlett JM, Horton K, Sue D, Thimothe J, Boor KJ, Wiedmann M (2001). Characterization and pathogenic potential of *Listeria monocytogenes* isolates from the smoked fish industry. *Appl. Environ. Microbiol.*, 67: 646-653.
- O'Grady J, Sedano-Balbás S, Maher M, Smith T, Barry T (2008). Rapid real-time PCR detection of *Listeria monocytogenes* in enriched food samples based on the *ssrA* gene, a novel diagnostic target. *Food Microbiol.*, 25: 75-84.
- Oravcová K, Kaclíková E, Krascenicsová K, Pangallo D, Brežná B, Siekel P, Kuchta T (2006). Detection and quantification of *Listeria monocytogenes* by 5'-nuclease polymerase chain reaction targeting the *actA* gene. *Lett. Appl. Microbiol.*, 42: 15-18.
- Oravcová K, Kuchta T, Kaclíková E (2007). A novel real-time PCR-based method for the detection of *Listeria monocytogenes* in food. *Lett. Appl. Microbiol.*, 45: 568-573.
- Pangallo D, Kaclíková E, Kuchta T, Drahovská H (2001). Detection of *Listeria monocytogenes* by polymerase chain reaction oriented to *inB* gene. *New Microbiol.*, 24: 333-339.
- Rantsiou K, Alessandria V, Urso R, Dolci P, Cocolin L (2008). Detection, quantification and vitality of *Listeria monocytogenes* in food as determined by quantitative PCR. *Intl. J. Food Microbiol.*, 121: 99-105.
- Rawsthorne H, Phister TG (2006). A real-time PCR assay for the enumeration and detection of *Zygosaccharomyces bailii* from wine and fruit juices. *Intl. J. Food Microbiol.*, 112: 1-7.
- Reichert-Schwillinsky F, Pin C, Dzieciol M, Wagner M, Hein I (2009). Stress- and growth rate-related differences between plate count and real-time PCR data during growth of *Listeria monocytogenes*. *Appl. Environ. Microbiol.*, 75: 2132-2138.
- Rodríguez-Lázaro D, Hernández M (2006). Isolation of *Listeria monocytogenes* DNA from meat products for quantitative detection by real-time PCR. *J Rapid Meth Autom Microbiol.*, 14: 395-404.
- Rodríguez-Lázaro D, Hernández M, Pla M (2004a). Simultaneous quantitative detection of *Listeria* spp. and *Listeria monocytogenes* using a duplex real-time PCR-based assay. *FEMS Microbiol Lett.*, 233: 257-267.
- Rodríguez-Lázaro D, Hernández M, Scotti M, Esteve T, Vázquez-Boland JA, Pla M (2004b). Quantitative Detection of *Listeria monocytogenes* and *Listeria innocua* by Real-Time PCR: Assessment of *hly*, *iap*, and *lin02483* Targets and AmpliFluor Technology. *Appl. Environ. Microbiol.*, 70: 1366-1377.
- Rodríguez-Lázaro D, Jofré A, Aymerich T, Hugas M, Pla M (2004c). Rapid quantitative detection of *Listeria monocytogenes* in meat products by real-time PCR. *Appl. Environ. Microbiol.*, 70: 6299-6301.
- Rodríguez-Lázaro D, Pla M, Scotti M, Monzó HJ, Vázquez-Boland JA (2005). A novel real-time PCR for *Listeria monocytogenes* that monitors analytical performance via an internal amplification control. *Appl. Environ. Microbiol.*, 71: 9008-9012.
- Rossmann P, Krassnig M, Wagner M, Hein I (2006). Detection of *Listeria monocytogenes* in food using a combined enrichment/real-time PCR method targeting the *prfA* gene. *Res. Microbiol.*, 157: 763-771.
- Skånseng B, Kaldhusdal M, Rudi K (2006). Comparison of chicken gut colonisation by the pathogens *Campylobacter jejuni* and *Clostridium perfringens* by real-time quantitative PCR. *Mol. Cell. Probes.*, 20: 269-279.

Ueda F, Anahara R, Yamada F, Mochizuki M, Ochiai Y, Hondo R (2005). Discrimination of *Listeria monocytogenes* contaminated commercial Japanese meats. Intl. J. Food Microbiol., 105: 455-462.

Zhang Y, Yeh E, Hall G, Cripe J, Bhagwat AA, Meng J (2007). Characterization of *Listeria monocytogenes* isolated from retail foods. Intl. J. Food Microbiol., 113: 47-53.